

Instructions for Use



Sperm Separation Device | ZMH0850

Important Information:

- Carefully adhere to recommended volumes for each step. Avoid over- or under-filling the device.
- Do not exceed 30 minute incubation time.
- Keep the device level during use – do not tip or rock.
- Device is single-use only and should be restricted to a single individual per device. It may not be reused.

Note on Incubation:


Good tissue practices necessitate matching media to incubation conditions. If using a bicarbonate-buffered media, incubate in a humidified, 37°C, gassed incubator. If using a HEPES-buffered media, incubate in a humidified, non-gassed incubator. If no incubator with humidity is available, add a 35mm dish of deionized or distilled water, uncovered, to the Petri dish containing the device before placing the covered dish with the device and the 35mm dish into the 37°C incubator.

Device Components:

- ZyMöt™ Multi (850µL) Sperm Separation Device
- Instructions for Use

Materials/Equipment Required, But Not Supplied:

- Sperm washing solution (media): bicarbonate- or HEPES-buffered media supplemented with 2-10% protein
- 37°C incubator
- >90mm Petri dish
- 1mL Luer-tip syringes (3) - Recommended: Norm-Ject #4010-200V0, Henke Sass Wolf
- Sperm-safe culture tube



PREPARATION

1. Gather your supplies and work on a clean surface.
2. Incubate semen sample at 37°C for 20-30 minutes to allow for liquefaction.
3. Carefully open the device package (Figure 1) without touching the device membrane.



Figure 1. Opened package and device.

DRAW SAMPLE

4. Use a 1mL syringe to slowly draw an 850µL aliquot of the liquefied semen specimen (sample). If there is insufficient sample volume, add media to bring volume to 850µL (Figure 2).

Note if using a frozen sample: Follow cryobank instructions when thawing. Dilute thawed sample 1:1 with media. Gently mix. Inject 850µL of the diluted sample into the device.



Figure 2. Slowly draw 850µL of the sample.

INJECT SAMPLE

5. Holding the device securely, carefully insert syringe into the device Inlet Port, applying gentle pressure to achieve a firm connection between syringe and device (Figure 3).




Figure 3. Firmly insert syringe into Inlet Port.

6. Apply slow and steady pressure to inject the sample (Figure 4).

Note: Be careful to avoid the formation of bubbles under the membrane.



Figure 4. Sample fully injected into device.

7. Prepare a fresh syringe with 750µL of media.
 - a) Prime the Outlet Port by injecting a small volume of media (approximately 50µL – Figure 5a), until the media travels through the channel to the membrane.

ADD MEDIA

b) Disconnect the syringe from the Outlet Port and apply the remaining media in the syringe to the surface of the upper membrane by dropping from approximately 2cm above the membrane (Figure 5b).

Completely cover the upper membrane with media, making sure media touches all the edges of the upper chamber and connects with the droplet of media that was used to prime the Outlet Port.

Note: Do not tilt the device to spread the media.



Figure 5a. Prime Outlet Port.




Figure 5b. Cover membrane surface.

INCUBATE SAMPLE

- Place device into a Petri dish and cover. Keep the ZyMöt device horizontal and covered at all times during the incubation. Incubate at 37°C for 30 minutes.

COLLECT SEPARATED SPERM

- Insert a fresh 1mL syringe into the Outlet Port, achieving a firm connection. Slowly aspirate a maximum of 500µL of the sperm-containing fluid (Figure 6).



Figure 6. Slowly aspirate a maximum of 500µL.

- Transfer the collected sample to an appropriate culture tube: a 4mL round bottom culture tube with a snap top or into the bottom of a 15mL conical tube. Tubes using HEPES-buffered media may be held on the benchtop or tightly capped in an incubator. Tubes using bicarbonate-buffered media should be stored in a CO₂ incubator with the lid loosely closed.

Tips, Warnings and Precautions:

- Caution: Federal law restricts this device to sale by or on the order of a physician.
- Device should be used only by properly trained operators.
- Practice universal precautions when handling human body fluids.
- Do not use if the packaging is damaged.
- Do not clean device with any solutions. Spraying or wiping with cleaning solutions can damage the device.

Device Description:

ZyMöt ICSI and ZyMöt Multi are sperm separation devices used to prepare motile sperm for assisted reproductive technology (ART) procedures. Both devices separate sperm based on motility. The ZyMöt ICSI and the ZyMöt Multi are sterile and single use only. The mechanism of action for both is separation of sperm based on motility within a microenvironment created by the micro channels of the ZyMöt ICSI or the micropores in the filter of the ZyMöt Multi. The primary difference between the devices is the processing volume. The ZyMöt ICSI has a processing volume of 2µL per micro channel. The ZyMöt Multi is manufactured in two (2) processing volumes, 850µL and 3mL.

The ZyMöt Multi (provided with 850µL and 3mL collection chambers) has an inlet port that communicates with the lower sample chamber. The sample chamber is separated from the upper collection chamber by a microporous filter. Untreated semen is added through the inlet port. After 30 minutes, the separated sperm are collected from the upper chamber through the outlet port.

Indications for Use:

The ZyMöt Multi (850µL) Sperm Separation Device is intended for preparing motile sperm from semen for use in the treatment of infertile couples by intracytoplasmic sperm injection (ICSI) and intrauterine insemination (IUI) procedures.

Note: To date, all studies and published clinical data regarding the use of ZyMöt Sperm Separation Devices have been focused on ICSI and IUI procedures. There have been no studies conducted, nor published data, on the use of ZyMöt devices for conventional insemination. ZyMöt Fertility is currently in the process of obtaining greater understanding and insight regarding the use of ZyMöt devices for conventional insemination. Consequently, we are requesting that clinicians limit their use of ZyMöt devices to ICSI and IUI cycles only.

Sterilization:

The sterilization method used for the ZyMöt devices is gamma radiation, at a dose level of 25kGy to 45kGy by the VD_{max}²⁵ method to meet a Sterility Assurance Level of 10⁻⁶.

Storage:

Store at 60°F - 77°F (15°C - 25°C).

Disposal:

Discard the used device and materials as medical waste.

Testing Performed for Devices Used in Assisted Reproduction:

Specific testing was performed for toxicity and functional screening appropriate for products used in assisted reproduction. As required by 21 CFR 884.6160, the following Special Controls were conducted (all tests were passed): human sperm survival assay (replacing the mouse embryo assay) and endotoxin testing.

Endotoxin Testing Results:

Using the Limulus Amebocyte Lysate (LAL) Analysis by the Gel-Clot Method, results were <0.0729 EU per device, which meets the acceptance level of ≤20 EU per device.

Human Sperm Survival Assay Results:

Using the Human Sperm Survival Assay, results were 96.2% for ZyMöt ICSI and 97.7% for ZyMöt Multi; both results meet the acceptance level of motility ≥80% of control at 24h after exposure for 30min. Note: The above results are from testing required prior to USFDA 510(k) clearance. These tests are conducted on each manufacturing lot of devices as part of the lot release program. A CoC can be provided upon request.

Manufactured for:

ZyMöt Fertility, a business unit of DxNow, Inc.
401 Professional Drive, Suite 130
Gaithersburg, MD 20879-3429 USA | zymotfertility.com

Manufactured by:

Natech Plastics, Inc.
85 Remington Blvd., Ronkonkoma, NY 11779 USA | natechplastics.com

Licenses, Patents and Trademarks:

DxNow, ZyMöt, ZyMöt Multi and ZyMöt ICSI are trademarks of DxNow, Inc. Devices are manufactured and sold under the terms of DxNow's exclusive worldwide patent license from The Brigham & Women's Hospital, Inc., Boston, Massachusetts, USA.

USA Patent US10422737B2; EU Patent EP2710139B1; Japan Patent JP6524082B2; Australia Patent AU2014353050B2. Additional USA and other international patents pending.

