

Certificate of Analysis

Product name: Oosafe® Disinfection wipes

Product code: OODW-70

Batch/LOT No.: 04712

Production date: 2020-11

Expiry date: 2023-11

Analysis / Method	Specification / Reference No.	Result	Completion date
Appearance / Visual	Clear liquid inside wipes	Clear liquid inside wipes	NA
Odor / Olfactive	Odourless	Odourless	NA
pH-Rate (2% active substance in water) / pH meter (DIN EN 1262)	7,0 – 9,0	8,2	NA
Density, 20 °C, gr/cm ³ / Density measuring device	1,001 – 1,010	1,004	NA
Quaternary Ammonium Compounds / QUAT test stripe	1,9 – 2,2 %	2 %	NA
Water		The rest	NA
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-5024-1220	Passed	25-12-2020
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4762-1220SPAR	Passed	27-12-2020

Quality Control according to ISO 13458:2016

GOosafe with SparMED!

Florencia Garcia
Quality Control Department
SparMED ApS
Date: 04-01-2021


