

## Certificate of Analysis

**Product ID:** OOSAFE® 35 MM DISH, HIGH WALL, Non-treated surface

**Product code:** OOPW-TF03

**LOT No.:** 07993

**Expiry date:** 2024-04

**Storage Conditions:** 20-30°C, dry room, no exposure to sunlight

**Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-2192-0719	Passed	26-07-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S3964-0719SPAR	Passed	24-07-2019
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E9467-0719SPAR	Passed	22-07-2019
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	NA	Passed	NA

Quality Control according to ISO 13485:2016

### GOosafe with SparMED!

Inga Sileikaite,  
Quality Control Department  
SparMED ApS

**Date: 29-07-2019**

