

Certificate of Analysis

Product ID: Oosafe® Andrology Tube

Product code: OOPW-AT10

LOT No.: 08036

Expiry date: 2024-11

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

| Analysis | Reference No. | Results | Completion date |
|--|----------------|---------|-----------------|
| Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours. | SPAR-3231-0120 | Passed | 17-01-2020 |
| Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven. | S4242-0120SPAR | Passed | 15-01-2020 |
| Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device. | E9894-0120SPAR | Passed | 13-01-2020 |
| Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy. | NA | Passed | NA |

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Inga Sileikaite,
Quality Control Department
SparMED ApS

Date: 20-01-2020

