

## Certificate of Analysis

**Product ID: Oosafe® Andrology Tube**

**Product code: OOPW-AT10**

**LOT No.: 08072**

**Expiry date: 2025-10**

**Storage Conditions: 20-30°C, dry room, no exposure to sunlight**

**Quality Assurance:**

| Analysis   | Reference No.  | Results | Completion date |
|--|----------------|---------|-----------------|
| Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours. | SPAR-5026-1220 | Passed  | 26-12-2020      |
| Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.  | S4763-1220SPAR | Passed  | 22-12-2020      |
| Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria $< 0.03$ EU/device.                                       | E0661-1220SPAR | Passed  | 21-12-2020      |
| Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.   | N/A            | Passed  | N/A             |

Quality Control according to ISO 13485:2016

### GOosafe with SparMED!

Florencia Garcia  
Quality Control Department  
SparMED ApS  
Date: 08-01-2021

