



Doc. No.: QC_4.14.2. Rev. Date: 01.04.2019

SparMED ApS
VAT no.: DK 30 89 85 75
Ryttermarken 2, 3520, Farum, Denmark
Phone: +45 39 40 25 03
e-mail: info@sparmed.dk - www.sparmed.dk

Certificate of Analysis

Product ID: Oosafe® 60mm dish with label area grip
Product code: OOPW-ST06
LOT No.: 08083
Expiry date: 2025-10
Storage Conditions: 20-30°C, dry room, no exposure to sunlight
Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-6605-0821	Passed	04-09-2021
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5287-0821SPAR	Passed	03-09-2021
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E1451-0821SPAR	Passed	01-09-2021
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Andrius Daugintis
Quality Control Department
SparMED ApS
Date: 01-10-2021

