

## Certificate of Analysis

**Product ID:** Oosafe® Pasteur Pipette, 3ml  
**Product code:** OOPW-PP03  
**LOT No.:** 08109  
**Expiry date:** 2026-03  
**Storage Conditions:** 20-30°C, dry room, no exposure to sunlight  
**Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-5670-0421	Passed	25-04-2021
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5020-0421SPAR	Passed	22-04-2021
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E1023-0421SPAR	Passed	23-04-2021
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

### GOosafe with SparMED!

Florencia Garcia  
 Quality Control Department  
 SparMED ApS  
 Date: 10-05-2021


