

Certificate of Analysis

Product ID: Oosafe® Center well Dish with two compartments**Product code: OOPW-CW05****LOT No.: 08112****Expiry date: 2026-03****Storage Conditions: 20-30°C, dry room, no exposure to sunlight****Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-5941-0621	Passed	13/06/2021
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5132-0621SPAR	Passed	11/06/2021
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E1176-0621SPAR	Passed	09/06/2021
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Zahraa Baker
Quality Control Department
SparMED ApS
Date: 12-07-2021

