



Esco Medical Technologies, UAB  
Draugystes g. 19, Kaunas, 51230 Lithuania  
Tel.: +370 686 79269  
Mail: [stasys.polimaitis@escolifesciences.com](mailto:stasys.polimaitis@escolifesciences.com) Web: [www.esco-medical.com](http://www.esco-medical.com)

## Certificate of conformity

Name: CultureCoin®

REF: 1821493

LOT: 061123

This certificate certifies that the product indicated above has been tested in accordance with the following test and meets its specifications.

### Test results:

Description tests	Pass/Fail
Visual inspection	Pass
Limulus Amebocyte Lysate (LAL) test	Pass
Mouse Embryo Assay (MEA) test	Pass

Irradiation carried out according ISO 11137:2015; ISO 13485:2016 and MDR, 2017/745 requirements. Measured absorbed dose was from 27,8 kGy to 34,9 kGy.

Certificate issued: 2023-11-28

Quality engineer,  
Gediminas Zubras

