



**CERTIFICATE OF CONFORMANCE**

**Issue Date:** 14 Jun 2024

Page 1 of 1

Description: MarScreen® IgG  
Product: 0006024  
Package size: 2 x 0.8 ml  
Lot: #052283  
Exp. Date: 2025-09-17

TEST PROCEDURE	SPECIFICATIONS	RESULTS
<b>APPEARANCE</b>		
<b>IgG Beads</b>		
Color	Blue beads in colorless liquid	PASSED
Dispersion	Evenly dispersed beads	PASSED
Volume	0.8 ml ± 0.1ml	PASSED
<b>Anti-IgG Serum</b>		
Color	Faint yellow to amber	PASSED
Clarity	Opaque	PASSED
Volume	0.8 ml ± 0.1ml	PASSED
<b>PERFORMANCE</b>		
<b>IgG Beads</b>		
	Slide agglutination within acceptable limits	PASSED
<b>Anti-IgG Serum</b>		
	Slide agglutination within acceptable limits	PASSED

*This document has been prepared by:*

Regulatory Affairs  
Alpha Tec Systems, Inc.

Location: 1311 SE Cardinal Court, Suite 170, Vancouver, WA 98683 USA  
Tel 360.260.2779 ♦ Fax 360.260.3277  
Email: [regulatory@AlphaTecSystems.com](mailto:regulatory@AlphaTecSystems.com) ♦ Web: [www.AlphaTecSystems.com](http://www.AlphaTecSystems.com)

Rev. A  
Effective Date: MAR 2024