



CERTIFICATE OF CONFORMANCE

Issue Date: 14 Jun 2024

Description: MarScreen® IgA
Product: 0006022
Package size: 2 x 0.8 ml
Lot: #052330
Exp. Date: 2025-09-17

Page 1 of 1

TEST PROCEDURE	SPECIFICATIONS	RESULTS
APPEARANCE		
IgA Beads		
Color	Blue beads in colorless liquid	PASSED
Dispersion	Evenly dispersed beads	PASSED
Volume	0.8 ml ± 0.1ml	PASSED
Anti-IgA Serum		
Color	Faint yellow to amber	PASSED
Clarity	Opaque	PASSED
Volume	0.8 ml ± 0.1ml	PASSED
PERFORMANCE		
IgA Beads		
	Slide agglutination within acceptable limits	PASSED
Anti-IgA Serum		
	Slide agglutination within acceptable limits	PASSED

This document has been prepared by:

Regulatory Affairs
Alpha Tec Systems, Inc.

Location: 1311 SE Cardinal Court, Suite 170, Vancouver, WA 98683 USA
Tel 360.260.2779 ♦ Fax 360.260.3277
Email: regulatory@AlphaTecSystems.com ♦ Web: www.AlphaTecSystems.com

Rev. A
Effective Date: MAR 2024