



CERTIFICATE OF CONFORMANCE

Issue Date: 14 Aug 2024

Description: MarScreen® IgG
Product: 0006024
Package size: 2 x 0.8 ml
Lot: #052619
Exp. Date: 2025-10-29

Page 1 of 1

TEST PROCEDURE

SPECIFICATIONS

RESULTS

APPEARANCE

IgG Beads

Color	Blue beads in colorless liquid	PASSED
Dispersion	Evenly dispersed beads	PASSED
Volume	0.8 ml ± 0.1ml	PASSED

Anti-IgG Serum

Color	Faint yellow to amber	PASSED
Clarity	Opaque	PASSED
Volume	0.8 ml ± 0.1ml	PASSED

PERFORMANCE

IgG Beads

Slide agglutination within acceptable limits	PASSED
--	--------

Anti-IgG Serum

Slide agglutination within acceptable limits	PASSED
--	--------

This document has been prepared by:

Regulatory Affairs
Alpha Tec Systems, Inc.

Location: 1311 SE Cardinal Court, Suite 170, Vancouver, WA 98683 USA
Tel 360.260.2779 ♦ Fax 360.260.3277
Email: regulatory@AlphaTecSystems.com ♦ Web: www.AlphaTecSystems.com

Rev. 1
Effective Date: MAR 2024