

Dok. / Ref.	F 8.2.6.7
Leidimo Nr. / Issue No.	1.0
Data / Date	28/06/2024

Company: Esco Medical Technologies, UAB

Address: Gamybos g. 2, Ramučiai, Kauno r., 54468 Lithuania

Name: CultureCoin®

REF: 1821493

LOT: 071024

This certificate certifies that the product indicated above has been tested in accordance with the following test and meets its specifications.

Test results:

Description tests	Criteria	Result	Pass/Fail	Ref. No.
Visual inspection	WI 7.5.1.3	No defects	Pass	F 7.5.1.5
Limulus Amebocyte Lysate (LAL) test	≤20 EU/device	<0,500 EU/device	Pass	LAL.023.639.2024
Mouse Embryo Assay (MEA) test	average blastocyst formation rate >80%	96,7%	Pass	MEA-2024'10.08-00

Irradiation carried out according ISO 11137:2015; ISO 13485:2016 and MDR, 2017/745 requirements. Measured absorbed dose was from 28,1 kGy to 35,6 kGy.

Certificate issued (14/10/2024):

Quality engineer
Gediminas Zubras

