



**CERTIFICATE OF CONFORMANCE**

**Issue Date:** 05 Dec 2024

Description: QC-Beads™  
Product: 0006020  
Package size: 2 x 4 ml  
Lot: #0017604  
Exp. Date: 2026-06-02

Page 1 of 1

<b>TEST PROCEDURE</b>	<b>SPECIFICATIONS</b>	<b>RESULTS</b>
<b>APPEARANCE</b>		
<b>Hi QC Beads</b>		
Color	Suspension of uniform white microspheres	PASSED
Volume	4 ml ± 0.25ml	PASSED
<b>Lo QC Beads</b>		
Color	Suspension of uniform white microspheres	PASSED
Volume	4 ml ± 0.25ml	PASSED
<b>PERFORMANCE</b>		
<b>Hi QC Beads</b>		
	34-46 million beads/ml with a 0.1-mm thick hemacytometer chamber	PASSED
<b>Lo QC Beads</b>		
	16-24 million beads/ml with a 0.1-mm thick hemacytometer chamber	PASSED

*This document has been prepared by:*

Regulatory Affairs  
Alpha Tec Systems, Inc.

Location: 1311 SE Cardinal Court, Suite 170, Vancouver, WA 98683 USA  
Tel 360.260.2779 ♦ Fax 360.260.3277  
Email: [regulatory@AlphaTecSystems.com](mailto:regulatory@AlphaTecSystems.com) ♦ Web: [www.AlphaTecSystems.com](http://www.AlphaTecSystems.com)

Rev. 4  
Effective Date: MAR 2024