

**CERTIFICATE OF CONFORMANCE****Issue Date:** 05 Feb 2025

Description: LeucoScreen®  
Product: 0006044  
Package size: 25-300 Determinations  
Lot: #0019361  
Exp. Date: 2026-01-07

Page 1 of 1

<b>TEST PROCEDURE</b>	<b>SPECIFICATIONS</b>	<b>RESULTS</b>
<b>Leuco Buffer</b>		
Color	Colorless	PASSED
Clarity	Clear	PASSED
pH	7.6 – 7.9	PASSED
Volume	7 ml ± 0.5 ml	PASSED
<b>DAB</b>		
Color	Dark Brown	PASSED
Clarity	Opaque	PASSED
Volume	1 ml ± 0.1 ml	PASSED
<b>Peroxide</b>		
Color	Colorless	PASSED
Clarity	Clear	PASSED
Volume	1 ml ± 0.1 ml	PASSED
<b>Leuco Fixative</b>		
Color	Colorless	PASSED
Clarity	Clear	PASSED
Volume	12 ml ± 0.5 ml	PASSED
<b>Peroxidase</b>		
Color	Colorless	PASSED
Clarity	Clear	PASSED
Volume	0.5 ml ± 0.1 ml	PASSED

*This document has been prepared by:*

Regulatory Affairs  
Alpha Tec Systems, Inc.

Location: 1311 SE Cardinal Court, Suite 170, Vancouver, WA 98683 USA  
Tel 360.260.2779 ♦ Fax 360.260.3277  
Email: [regulatory@AlphaTecSystems.com](mailto:regulatory@AlphaTecSystems.com) ♦ Web: [www.AlphaTecSystems.com](http://www.AlphaTecSystems.com)

Rev. 4  
Effective Date: MAR 2024