



CERTIFICATE OF CONFORMANCE

Issue Date: 26 Feb 2025

Description: MarScreen® IgG
Product: 0006024
Package size: 2 x 0.8 ml
Lot: #0020118
Exp. Date: 2026-07-21

Page 1 of 1

TEST PROCEDURE	SPECIFICATIONS	RESULTS
APPEARANCE		
IgG Beads		
Color	Blue beads in colorless liquid	PASSED
Dispersion	Evenly dispersed beads	PASSED
Volume	0.8 ml ± 0.1 ml	PASSED
Anti-IgG Serum		
Color	Faint yellow to amber	PASSED
Clarity	Opaque	PASSED
Volume	0.8 ml ± 0.1 ml	PASSED
PERFORMANCE		
IgG Beads	Slide agglutination within acceptable limits	PASSED
Anti-IgG Serum	Slide agglutination within acceptable limits	PASSED

This document has been prepared by:

Regulatory Affairs
Alpha Tec Systems, Inc.

Location: 1311 SE Cardinal Court, Suite 170, Vancouver, WA 98683 USA
Tel 360.260.2779 ♦ Fax 360.260.3277
Email: regulatory@AlphaTecSystems.com ♦ Web: www.AlphaTecSystems.com

Effective Date: MAR 2024