



CERTIFICATE OF CONFORMANCE

Issue Date: 13 Mar 2025

Description: LeucoScreen®
Product: 0006044
Package size: 25-300 Determinations
Lot: #0020481
Exp. Date: 2026-01-07

Page 1 of 1

TEST PROCEDURE	SPECIFICATIONS	RESULTS
Leuco Buffer		
Color	Colorless	PASSED
Clarity	Clear	PASSED
pH	7.6 – 7.9	PASSED
Volume	7 ml ± 0.5 ml	PASSED
DAB		
Color	Dark Brown	PASSED
Clarity	Opaque	PASSED
Volume	1 ml ± 0.1 ml	PASSED
Peroxide		
Color	Colorless	PASSED
Clarity	Clear	PASSED
Volume	1 ml ± 0.1 ml	PASSED
Leuco Fixative		
Color	Colorless	PASSED
Clarity	Clear	PASSED
Volume	12 ml ± 0.5 ml	PASSED
Peroxidase		
Color	Colorless	PASSED
Clarity	Clear	PASSED
Volume	0.5 ml ± 0.1 ml	PASSED

This document has been prepared by:

Regulatory Affairs
Alpha Tec Systems, Inc.

Location: 1311 SE Cardinal Court, Suite 170, Vancouver, WA 98683 USA
Tel 360.260.2779 ♦ Fax 360.260.3277
Email: regulatory@AlphaTecSystems.com ♦ Web: www.AlphaTecSystems.com

Rev. 1
Effective Date: MAR 2024